



King's Research Portal

DOI: 10.2967/jnumed.116.181032

Document Version Peer reviewed version

Link to publication record in King's Research Portal

Citation for published version (APA):

Grootendorst, M. R., Cariati, M., Pinder, S. E., Kothari, A., Douek, M., Kovacs, T., Hamed, H., Pawa, A., Nimmo, F., Owen, J., Ramalingam, V., Sethi, S., Mistry, S., Vyas, K., Tuch, D. S., Britten, A., Van Hemelrijck, M., Cook, G. J., Sibley-Allen, C., ... Purushotham, A. (2017). Intraoperative assessment of tumor resection margins in breast-conserving surgery using ¹⁸F-FDG cerenkov luminescence imaging: A first-in-human feasibility study. *Journal of Nuclear Medicine*, *58*(6), 891-898. https://doi.org/10.2967/jnumed.116.181032

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
You may not further distribute the material or use it for any profit-making activity or commercial gain
You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Intraoperative Assessment of Tumor Resection Margins in Breast-Conserving Surgery using ¹⁸F-FDG Cerenkov Luminescence Imaging – A First-in-Human Feasibility Study

Maarten R. Grootendorst^{1,2} Massimiliano Cariati^{1,2} Sarah E. Pinder¹ Ashutosh Kothari² Michael Douek^{1,2} Tibor Kovacs² Hisham Hamed² Amit Pawa³ Fiona Nimmo⁴ Julie Owen¹ Vernie Ramalingam² Sweta Sethi² Sanjay Mistry² Kunal Vyas⁵ David S. Tuch⁶ Alan Britten⁷ Mieke Van Hemelrijck¹ Gary J. Cook⁸ Chris Sibley-Allen⁹ Sarah Allen⁹ Arnie Purushotham^{1,2}

Author affiliations:

- ¹ King's College London, Division of Cancer Studies, London, UK;
- ² Department of Breast Surgery, Guy's and St Thomas' NHS Foundation Trust, London, UK;
- ³ Anesthetic Department, Guy's and St Thomas' NHS Foundation Trust, London, UK;
- ⁴ Day Surgery Unit, Guy's and St Thomas' NHS Foundation Trust, London, UK;
- ⁵ Sagentia Ltd., Cambridge, UK;
- ⁶ Lightpoint Medical Ltd., Chesham, UK;
- ⁷ Medical Physics Department, St George's Hospital, London, UK;
- ⁸ King's College London, Division of Imaging Sciences and Biomedical Engineering, London, UK
- ⁹ Department of Nuclear Medicine, Guy's and St Thomas' NHS Foundation Trust, London, UK;

Conflicts of interest: David S. Tuch is an employee of and a shareholder in Lightpoint Medical Ltd. The other authors have no conflicts of interest to disclose.

Corresponding author

Name: Arnie Purushotham

Address: Department of Research Oncology, 3rd Floor Bermondsey Wing, Guy's Hospital, SE1 9RT, London

Phone: +44 (0) 207 188 3027

Email: arniepurushotham@gmail.com

First author

Name: Maarten Grootendorst (PhD student)

Address: Department of Research Oncology, 3rd Floor Bermondsey Wing, Guy's Hospital, SE1 9RT, London

Phone: +44 (0) 75 188 92692

Email: maarten.grootendorst@kcl.ac.uk

Word count: 6922

Source of Funding: Supported by funding from Innovate UK, Cancer Research UK King's Health Partners Experimental Cancer Medicine Centre, Guy's and St Thomas' Charity, and the National Institute for Health Research (NIHR) Biomedical Research Centre at Guy's and St Thomas' NHS Foundation Trust and King's College London. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health.

ABSTRACT

Rationale: In early-stage breast cancer, the primary treatment option for the majority of women is breast-conserving surgery (BCS). There is a clear need for more accurate techniques to assess resection margins intraoperatively, as on average 20% of patients require further surgery to achieve clear margins. Cerenkov luminescence imaging (CLI) combines optical and molecular imaging by detecting light emitted by 18F-fluorodeoxyglucose (¹⁸F-FDG). Its high-resolution and small size imaging equipment makes CLI a promising technology for intraoperative margin assessment. A first-in-human study was conducted to evaluate the feasibility of ¹⁸F-FDG CLI for intraoperative assessment of tumor margins in BCS.

Methods: Twenty-two patients with invasive breast cancer received 5 MBq/kg ¹⁸F-FDG 45-60 min prior to surgery. Sentinel lymph node biopsy (SLNB) was performed using an increased technetium-99m (^{99m}Tc)-nanocolloid activity of 150 MBq to facilitate nodal detection against the gamma-probe background signal (cross-talk) from ¹⁸F-FDG. The cross-talk and ^{99m}Tc dose required was evaluated in two lead-in studies. Immediately after excision, specimens were imaged intraoperatively in an investigational CLI imaging system. The first 10 patients were used to optimize the imaging protocol; the remaining 12 patients were included in the analysis dataset. CLI images from incised BCS specimens were analyzed postoperatively by two surgeons blinded to the histopathology results, and mean radiance and margin distance were measured. Agreement between margin distance on CLI and histopathology was assessed. Radiation doses to staff were measured.

Results: Ten of the 12 patients had an elevated tumor radiance on CLI. Mean radiance and tumor-tobackground ratio were 560 \pm 160 photons/s/cm²/sr and 2.41 \pm 0.54, respectively. All 15 assessable margins were clear on CLI and histopathology. Agreement in margin distance and inter-rater agreement was good (κ = 0.81 and 0.912, respectively). Sentinel lymph nodes (SLNs) were successfully detected in all patients. Radiation dose to staff was low; surgeons received a mean dose of 34 \pm 15 µSv per procedure. **Conclusions:** Intraoperative ¹⁸F-FDG CLI is a promising, low-risk technique for intraoperative assessment of tumor margins in breast-conserving surgery. A randomized controlled trial will evaluate the impact of this technique on re-excision rates.

Key words: Cerenkov luminescence imaging; breast-conserving surgery; tumor margins; ¹⁸F-FDG; sentinel lymph node biopsy

INTRODUCTION

In early-stage breast cancer, the primary treatment option for the majority of women is BCS by wide local excision (WLE) of the tumor. WLE often fails to achieve clear surgical margins, and on average 20% of patients who undergo BCS will require repeat surgery to achieve clear margins (*1*) (although this may vary since there is no global agreement of the definition of 'clear margins'). Re-operations potentially have several negative consequences including delayed commencement of adjuvant therapy, worse cosmesis, increased patient anxiety and costs (*2*,*3*).

There have been several attempts to assess surgical margins intraoperatively in order to reduce breast cancer re-operation rates post-WLE (1). Techniques evaluated to date include specimen radiography, intraoperative ultrasound, touch imprint cytology, frozen section, and radiofrequency spectroscopy. However, these all have limitations in terms of adequate performance, practicality and/or cost-effectiveness (1). Experimental methods evaluated include Raman spectroscopy, ambient mass spectrometry, optical coherence tomography, diffuse reflectance spectroscopy, confocal microscopy, and (targeted) fluorescence imaging (1). Each of these techniques have unique limitations, and the diagnostic performance remains to be evaluated in large-scale studies.

Positron emission tomography (PET) using ¹⁸F-FDG is a powerful tool for *in vivo* imaging of breast cancer. While whole-body PET has limited diagnostic sensitivity for primary breast cancer, high-resolution PET imaging with positron emission mammography has shown high sensitivity (92-96%) and specificity (84-91%) for breast cancer diagnosis (*4-6*). Intraoperative high-resolution imaging of ¹⁸F-FDG could therefore provide a powerful tool for surgical guidance. However, intraoperative PET is impractical due to the large size and expense of a PET scanner and PET's low spatial resolution. Development of a compact, high-resolution, intraoperative PET scanner could address these limitations.

Recently, it has been discovered that PET imaging agents emit optical photons via a phenomenon called Cerenkov luminescence (7). Cerenkov photons are generated by positrons travelling at super-relativistic speeds in tissue. Optical imaging of Cerenkov photons emitted by PET agents is an emerging imaging modality called CLI. CLI combines high diagnostic performance and clinical translatability of PET imaging with high spatial resolution and compactness of optical cameras,

thus making it a promising technology for intraoperative margin assessment in breast cancer surgery (8).

In this first-in-human clinical trial we evaluated the feasibility, safety, and preliminary performance of ¹⁸F-FDG CLI using a novel intraoperative CLI camera to assess tumor margin status in breast cancer patients undergoing WLE with SLNB or with axillary lymph node dissection (ALND).

MATERIALS AND METHODS

Intraoperative ¹⁸F-FDG CLI in breast-conserving surgery

Patient recruitment and patient preparation on the day of surgery. Research Ethics Committee approval was obtained prior to patient recruitment (ClinicalTrials.gov identifier NCT02037269). Between June 2014 and February 2016, patients with histologically confirmed invasive breast cancer on core biopsy with or without associated ductal carcinoma in situ (DCIS), due to undergo primary BCS and SLNB or ALND, were recruited at Guy's Hospital in London after written informed consent was obtained. Exclusion criteria were: age <30 years, previous surgery or radiotherapy to the ipsilateral breast in the preceding two years, neoadjuvant systemic therapy, pregnancy or lactation, blood glucose level of \geq 12 mmol/l on the day of surgery and known hypersensitivity to ¹⁸F-FDG. Females of childbearing age required a negative pregnancy test (by beta HCG qualitative analysis), history of surgical sterilization, or history of amenorrhea in the past twelve months.

On the day of surgery patients scheduled to undergo SLNB received a periareolar intradermal injection of 150 MBq ^{99m}Tc-albumin-nanocolloid (Nanocoll[™], GE Healthcare, UK). The increased ^{99m}Tc activity of 150 MBq was calculated based on the results from two lead-in cross-talk studies (Supplemental Material). Patients were then injected intravenously with 5 MBq/kg ¹⁸F-FDG (up to a maximum of 300 MBq) and typically 45-60 min post-¹⁸F-FDG injection were taken to the operating theatre.

Surgery and intraoperative specimen radiography. Following induction of anesthesia, patients due to undergo SLNB received a periareolar subdermal injection of 2ml Patent Blue V and 3ml of normal saline. To minimize radiation exposure to theatre staff by reducing the time spent in close proximity to the patient, a standard breast operating set was pre-arranged on a sterile tray. Surgery to the breast

was performed ahead of SLNB/ALND to minimize signal intensity reduction from radiotracer decay in the time between ¹⁸F-FDG injection and CLI imaging. The WLE specimen was excised using monopolar diathermy (Valleylab Force FXTM electrosurgical generator with HCP-01 Skintact surgical pencil). The excised specimen was orientated with sutures and metal surgical clips as per local protocol.

Post-excision WLE specimens were x-rayed intraoperatively (Faxitron Bioptics, USA), and excision of cavity shave margins was performed if the tumor was deemed to be close to the edge of the specimen on radiography.

Following excision of the WLE specimen, SLNB or ALND was performed. For SLNB a Europrobe 3 gamma probe with a high-energy collimator was used (Eurorad SA, France). SLNs were defined as nodes that were radioactive, blue, or palpable (9). The number of excised SLNs, the *ex vivo* SLN gamma probe signal (counts per second), and the presence of blue nodal discoloration were recorded. Upon completion of the procedure the gamma probe background signal in the axilla was measured.

Intraoperative CLI of WLE specimens and lymph nodes. Following specimen radiography, CLI imaging of the WLE specimen was performed using an investigational intraoperative CLI imaging system (Lightpoint Medical Ltd, UK). This system consists of a custom-built light-tight dark box containing two optical pathways: one for CLI and one for white-light imaging for anatomic reference (Fig. 1A). The CLI imaging pathway includes a fast f/.95 lens and a reflex mirror to fold the optical pathway into an electron-multiplying charge coupled device (EMCCD) camera. The field-of-view of the CLI camera is 8×8 cm, and the acquisition matrix is 512×512 to give a pixel resolution of 156.25 µm. The EMCCD is thermoelectrically-cooled to -80°C and radiation-shielded with lead to prevent annihilation photons from scintillating in the EMCCD chip, i.e. "gamma strikes". The white-light imaging pathway provides a photographic reference image using a standard complementary metal-oxide-semiconductor camera.

The WLE specimen was positioned on a specimen table (Fig. 1B), the margin of interest placed in the center of the field-of-view by using the surgical sutures to guide orientation, and subsequently imaged.

Following intact WLE specimen imaging, the surface of the specimen was immediately inked intraoperatively in order to preserve its orientation for histopathological analysis (Supplemental Material Fig. 1A). Six distinct ink colors (Davidson[®], Bradley Products Inc., USA) were applied to the six margins. The inked specimen was then incised through the posterior margin to visualize the primary tumor and tumor margins, and the incised WLE specimen was imaged (Supplemental Material Figs. 1B and 1C). In one patient sequential image acquisition over a 50-minute time-period was performed to determine the half-life ($t_{1/2}$) of the radiance observed in the tumor.

The first 10 patients were included in the optimization dataset and the remaining 12 patients in the analysis dataset. In the first 10 patients, the image acquisition protocol was optimized by testing different image acquisition times (100, 300, 400 sec) and pixel binning settings (2x2, 4x4, 8x8). A 300-sec acquisition time and 8x8 pixel binning was found to provide sufficient sensitivity for tumor detection and acceptable spatial resolution (1.25 mm) within a time-window feasible for intraoperative use, and these setting were used in the remaining 12 patients included in the analysis dataset. Upon completion of WLE CLI, the activity of the WLE specimen was estimated using a scintillation monitor (Type 41/44A, ThermoScientific, USA) or handheld radiation spectrometer (Raymon10 GR1, Kromek PLC). SLNs were also imaged intraoperatively with CLI using the same imaging settings.

After imaging was completed, WLE specimens were sent for histopathological analysis as per standard practice.

Radiation safety monitoring. Radiation safety monitoring was performed to ensure safe working practices were maintained and that work was compliant with UK legislation regarding ionizing radiation (*10-12*). Prior to commencing the study all staff received training to become familiar with radiation control procedures, occupational risks, and learned how to minimize exposure without compromising patient care. Staff members were issued with electronic personal radiation dose monitors (PDM-112 and PDM-122, Hitachi-Aloka Medical Ltd., Japan) for the body, and thermo-stimulated luminescent ring dosimeters for extremities (Landauer, UK). Radiation contamination monitoring of staff, rooms, equipment and waste was carried out after each procedure using a scintillation monitor (Type 41/44A, Series 300 mini-monitor, ThermoScientific, USA). As ^{99m}Tc has a longer half-life (6.02 hours) than ¹⁸F (110 min) the radioactive waste storage requirements for CLI procedures are similar to standard SLNB

procedures. The time taken for the various stages of the procedure, i.e. from induction of anesthesia to recovery, were recorded.

Histopathology. Histopathological analysis was performed as per UK National guidelines: the WLE specimen was sliced at 2 mm intervals, and representative sections of the tumor and all 6 relevant margins were selected by the pathologist, processed, paraffin wax embedded and 3 to 4 micron sections were cut and stained with hematoxylin and eosin. Microscopic margin distance measurements were performed by a Consultant breast pathologist (S.E.P.). Microscopic invasive tumor size and whole tumor size (including DCIS extending from the main invasive mass) were also measured. Positive margins were defined as invasive cancer or DCIS <1 mm from the specimen surface. The histological margin distances were reported in increments of 1 mm, but margins more than 5 mm were reported as >5 mm. The pathologist was blinded to the interpretation of the CLI images.

Image analysis. All CLI and radiography images were analyzed postoperatively in order to provide a controlled and standardized analysis environment. Measurements of the mean radiance (photons/s/cm²/sr) were performed by drawing region-of-interests on the unprocessed CLI images. Region-of-interests were selected in areas showing increased signal intensity ('tumor') and no increased signal ('tissue background'). Tumor-to-background ratios (TBRs) were calculated. Gamma strikes were excluded from region-of-interest analysis. The tumor radiance from the sequential incised WLE images was fit to a monoexponential, to determine the radiance half-life.

Assessment of margin status on CLI was performed on the incised WLE specimen images. The analysis was done independently by two experienced breast surgeons (AP and AK), and performed prior to analysis of the radiography images to prevent potential confirmation bias from *a priori* knowledge of the radiological margin status. Prior to analysis CLI images were processed by applying a median filter (filter size range 5 - 10, filter threshold range 10 - 15) and Gaussian filter (filter width: 1, filter threshold: 0.5). A stronger Gaussian filter (filter width: 4 or 5) was applied to images with a low TBR to increase the visibility of the tumor. The preoperative diagnostic information that would typically be available to the surgeon was provided including patient age, clinical, mammographic and ultrasound tumor size, screen detected (Y/N), and histological tumor type, grade

and receptor status on core biopsy. Per patient, a color image containing information on specimen orientation was shown together with a grey-scale image and Cerenkov image. All images were displayed on a standard computer monitor (23", 1920 x 1080 pixels, 250 cd/m² luminance). The grey-scale image was overlaid with the Cerenkov signal to provide a fused image containing both functional and anatomical information. The leveling was set using the software's default leveling, and manually adjusted based on the surgeon's clinical judgment. Both surgeons then independently reported whether an elevated radiance from the tumor could be identified on CLI; in patients displaying an elevated tumor radiance the margin distance of the margins visible in the image was measured using the ruler function in the imaging software (Mirada XD3, Mirada Medical, Oxford, UK). The total time required to complete margin assessment was approximately 2 min per patient. As an exploratory outcome measure, tumor size was also measured. Upon completion of the measurements surgeons were asked whether, given the CLI image, they would have performed a cavity shaving had the image been available at the time of surgery. Surgeons also scored image quality on a 5-point Likert scale: 1 = very poor – image not interpretable, 2 = poor but interpretable, 3 = fair, 4 = good, 5 = very good.

Following CLI image analysis, specimen radiography image analysis was performed on a Coronis 3MP screen (20.8", 1536 x 2048 pixels, 500 cd/m² luminance) using standard GE PACS imaging software. Surgeons were presented with the same preoperative diagnostic information, but the images were shown in a different order to avoid potential sequential bias. The number of surgical marker clips was noted, and the reliability of specimen orientation assessed. If the orientation was considered reliable, the margin distance and tumor size on radiography was measured. Whether an additional cavity shaving would have been performed based on the radiography image was also noted.

The final histopathology results of the surgically excised tissue were not available at the time of CLI and radiography image analysis, and could therefore not bias the surgeon's assessment.

Statistics. Weighted Kappa coefficients were calculated to assess the agreement in margin distance between CLI and definitive histopathology, and to assess the inter-rater agreement between surgeons ('irr' package, version 3.2.2, R statistical software). A kappa coefficient (κ) greater than 0.75 was considered good agreement (*13*). Agreement between histological tumor size and tumor size on

CLI and radiography respectively was assessed by calculating the mean difference in tumor size ± std, and intraclass correlation coefficients (ICC) (SPSS® version 23.0; IBM, Chicago).

RESULTS

Intraoperative imaging of WLE specimens

A total of 22 patients were included in the study. The CLI results and postoperative histopathology results from the 12 patients included in the analysis dataset are shown in Table 1. The mean administered ¹⁸F-FDG activity was 295 ± 18 MBq (range 259 – 325). The mean time between ¹⁸F-FDG injection and WLE excision was 86 ± 26 min (range 50 – 146), and the mean time between ¹⁸F-FDG injection and commencement of CLI image acquisition was 118 ± 26 min (range 88 – 180).

Tumor margin assessment was performed on incised WLE specimen images to allow for visualization of the tumor extent and to avoid image artifacts created by the monopolar diathermy. Ten of the 12 patients in the analysis dataset had elevated tumor radiance on CLI (Table 1). Mean tumor radiance and TBR in these 10 patients was 560 ± 160 photons/s/cm²/sr (range 308 - 871) and 2.41 ± 0.54 (range 1.63 - 3.22). The half-life of the tumor radiance was 115.5 min, which is consistent with the 109.8-minute half-life of ¹⁸F. This concordance supports that the detected tumor radiance is Cerenkov luminescence from ¹⁸F-FDG. Mean radioactivity in the WLE specimen at the time of CLI imaging was 90 ± 48 kBq in patients with an elevated radiance; in the 2 patients without an elevated radiance radioactivity was 14 kBq and 19 kBq, respectively.

In the 10 patients with elevated tumor radiance, a total of 60 margins could be assessed histologically, 26 margins were evaluable on specimen radiography, and 15 margins were assessable on CLI. Of the 45 histological margins that were not evaluable on CLI, 40 were not in the field-of-view of the CLI image, and 5 could not be assessed due to migration of the specimen orientation ink onto the margin edge, preventing optical margin interrogation. Eighteen of the 60 histological margins were not assessable on specimen radiography due to the inability to reliably orientate the specimen on the radiography image, and 16 margins were not in the image field-of-view.

The margin distance from the 15 margins as measured on CLI and histopathology is shown in Table 1. Two margins measured between 1 and 5 mm on CLI and histopathology (Figs. 2 and 3); the remaining 13 margins were >5 mm by both modalities. There was good agreement between the histological margin distance and the margin distance on CLI as measured by both surgeon 1 and

surgeon 2, respectively (κ surgeon 1 = 0.76, κ surgeon 2 = 0.86). The agreement in margin distance between surgeons was also good (κ = 0.91).

Five margins could be assessed on both CLI and specimen radiography, and all were >5 mm on both modalities, as well as histologically. An example of a CLI, radiography and histopathology image from a patient with >5 mm resection margin widths is shown in Fig. 4.

Two patients (17%) had a positive margin on postoperative histopathological analysis; both were medial margins with DCIS <1 mm distant. These margins were not visible in the CLI image as specimen incision had only exposed the superior, inferior and posterior margins; the medial margin could therefore not be assessed.

In 8 of the 10 patients tumor size could be measured on CLI, and compared to histopathology: the agreement is shown in Table 1. In 2 patients the orientation inks prevented measurement of tumor size on CLI. Invasive tumor size showed excellent agreement; mean difference for both surgeons combined was -0.84 ± 2.8 mm. ICC was 0.84 and 0.81 for surgeons 1 and 2, respectively. Whole tumor size was underestimated on CLI; mean difference for both surgeons combined was -4.7 ± 5.0 mm. ICC was 0.65 and 0.69 for surgeons 1 and 2, respectively. Inter-rater agreement between surgeons was excellent (ICC = 0.97).

The agreement between invasive tumor size on histopathology and on radiography was good; mean difference for both surgeons combined was 1.0 ± 3.1 mm. ICC was 0.56 and 0.58 for surgeons 1 and 2, respectively. Whole tumor size was underestimated on radiography; mean difference for both surgeons combined was -5.2 ± 8.9 mm.

CLI image quality in the 10 patients with successful CLI was scored as 4.3 (range 4 - 5) by both surgeons.

Sentinel lymph node detection and ¹⁸F-FDG Cerenkov lymph node imaging

SLNB was performed in 21 of the 22 patients; 1 patient underwent an ALND. SLNs were successfully identified in all 21 patients. A total of 43 SLNs were removed. The average number of SLNs per patient was 2 (range 1 - 4). Two of the 21 SLNB patients had macrometastatic SLNs.

The mean gamma probe signal of the 'hottest' SLN per patient was 4991 \pm 2521 counts per second (range 8500 – 170). The mean gamma probe signal of the 'second hottest' SLN was 2505 \pm

2632 counts per second (range 7368 – 50). Mean axillary background signal, measured in 13 patients, was 192 \pm 70 counts per second (range 55 – 270). This signal is lower than the ¹⁸F-FDG gamma probe cross-talk measured in the lead-in study (Supplemental Material), and is mainly due to the longer time between ¹⁸F-FDG injection and SLNB (mean 93 \pm 34 min). A total of 7 nodes had a gamma probe signal below the background signal; 6 of these were blue. This indicates the importance of using the combined technique of radioisotope and blue dye in ¹⁸F-FDG CLI guided breast surgery, as low-uptake nodes may be missed if gamma probe detection is used alone.

All SLNB procedures were performed with the monopolar diathermy device and due to the observed image artifact from diathermy on CLI, the SLN images were uninterpretable.

Radiation dose to staff

A summary of the whole body effective radiation dose to primary personnel from all 22 procedures is shown in Table 2. Surgeons received the highest mean and maximum dose of 34 μ Sv and 74 μ Sv respectively. Mean duration of surgery was 39 ± 11 min (range 21 – 61) during which the surgeon was generally <0.5 meters from the patient. Mean radiation dose to the left and right hand of the surgeon was 126 ± 95 μ Sv (0 – 250) and 78 ± 75 μ Sv (0 – 200), respectively. Mean and maximum radiation dose received by the anesthetist standing at approximately 1 meter from the patient, with closer patient contact at the time of induction of anesthesia and at the end of the procedure, was 11 μ Sv and 18 μ Sv, respectively. Surgical equipment had low levels of radioactive contamination, which was undetectable 1-3 days later. No staff members were found to be contaminated with radioactivity after the procedures.

DISCUSSION

This first-in-human study evaluated the feasibility of intraoperative ¹⁸F-FDG CLI for assessing tumor margin status in patients with invasive breast cancer undergoing BCS, and SLNB or ALND. Tumor margin assessment on CLI could be performed in 10 of the 12 patients in the analysis dataset, and there was strong agreement between CLI and definitive histopathology on margin width. An exploratory outcome measure assessed the correlation between tumor size on CLI and histopathology; the size on CLI and histopathology correlated well for invasive cancer, while whole tumor size (invasive with associated DCIS) was underestimated on CLI. Results from the radiation

monitoring program demonstrated that the procedure can be carried out safely while maintaining low radiation exposures to the staff involved.

In 2 patients margin assessment could not be performed because the tumors did not display elevated radiance on CLI. The absence of signal in these patients is probably due to the small tumor size, a factor known to be associated with lower ¹⁸F-FDG uptake (*14*), and the late time points at which these tumors were imaged (135 min and 180 min post ¹⁸F-FDG injection; the first and third longest injection-imaging time of all patients). Unsuccessful CLI imaging due to the absence of a detectable tumor signal highlights the importance of ongoing developments focused on improving detection sensitivity of camera systems to aid detection of tumors with low ¹⁸F-FDG uptake including lower grade tumors and DCIS (*4*).

Since its discovery in 2009, CLI has rapidly emerged as a powerful technique for cancer imaging. CLI is readily translatable to the clinic due to existing regulatory approval and widespread availability of PET imaging agents (*15*). In contrast, targeted fluorescence imaging requires prohibitive clinical development times and capital investment for regulatory and reimbursement approval of novel imaging drugs (*16*). Three clinical pilot studies of CLI have been published to date. These have focused on the use of CLI to image radiopharmaceutical uptake in the thyroid, CLI for non-invasive detection of nodal disease, and Cerenkov luminescence endoscopy to aid detection of cancerous lesions in the GI tract (*17-19*). To our knowledge, this is the first report of intraoperative CLI. Its high-resolution, small size imaging equipment and minute-scale image acquisition (5 min) and image analysis (~2 min) times, make CLI of particular interest for image-guided surgery. The feasibility of intraoperative CLI as shown in this study in combination with the wide applicability of ¹⁸F-FDG across a range of solid cancers provides a stepping stone for clinical evaluation of this technology in other cancer types.

The low radiation exposure to staff found in this study is in accordance with previously reported exposure levels from ¹⁸F-FDG guided breast surgery procedures (20,21), and comparable to the radiation dose reported for interventional cardiology procedures (1 – 50 μ Sv) (22). The number of ¹⁸F-FDG CLI-guided BCS procedures that could be performed in a routine clinical setting depends on the occupational limits on radiation exposure per country (Table 2). In the UK and US the occupational annual dose limit is 20 mSv (23) and 50 mSv (24), respectively. Good practice would dictate that the radiation exposure from a procedure should be kept As Low As Reasonable Achievable, i.e. well

below the dose limits. In practice in the UK, if a worker is likely to receive annually more than 6mSv they would be designated a classified worker, necessitating annual medical surveillance and longer term record keeping of their radiation exposure.

Image artifacts on CLI from tissue excised with the monopolar diathermy device prevented tumor margin assessment on intact WLE specimens and assessment of SLNs. Although the source of this 'false-signal' is not yet fully understood, current evidence from pre-clinical experiments points towards long-lived, thermally-induced chemiluminescence (*25*). Since the emission seems to be related to temperature, which can reach up to 250°C at the tip of the diathermy device, electrosurgical devices that operate at much lower temperatures are currently being tested (*26*). In addition to potentially facilitating margin assessment on intact WLE specimens, an advantage of low-temperature devices over monopolar diathermy is the reduced collateral tissue damage, which could also improve the accuracy of assessing tumor resection margins on histopathology (*27*).

Although CLI imaging of incised WLE specimens is feasible for assessing tumor margin status, this approach has some limitations over margin assessment on intact specimens. Firstly, migration of the wet pathology ink onto the margin edge immediately after specimen incision hinders margin interpretation with CLI. Methods to accelerate drying of inks by applying acetic acid to the painted tissue or by using fast drying inks may be solutions to this problem, but this has not been tested in this study. Secondly, in our institution specimen incision could only be performed through the posterior margin to ensure accurate postoperative histological assessment of radial margins. Consequently, only a limited number of margins could be assessed with CLI imaging per patient, and two histologically positive margins that were not visible in the CLI image were therefore missed. In order to assess more margins per patient specimen incision may be performed in multiple planes, but good communication between surgeons and pathologists is paramount in order to not compromise patient care.

A randomized, controlled, multi-center clinical study is scheduled to commence in late 2016 to evaluate the effect of intraoperative ¹⁸F-FDG CLI on re-operation rate and Quality of Life in BCS (ClinicalTrials.gov identifier NCT02666079). The study will run across an anticipated 8 study sites in the UK and Germany and use the CE-marked LightPath[™] Imaging System (Lightpoint Medical Ltd, UK). The smaller field-of-view of 6 x 6 cm and improved imaging software may provide substantial improvements in sensitivity over the investigational CLI camera used in the present study. By

analyzing larger subgroups of patients with a range of tumor types (including DCIS), size, histological grades and hormone receptor status, further insight should be obtained into which breast-cancer patient populations may most benefit from CLI-guided surgery.

CONCLUSION

Intraoperative ¹⁸F-FDG CLI in BCS for invasive breast carcinoma is a promising and low-risk procedure. CLI imaging of incised WLE specimens provides high-resolution functional information that allows surgeons to accurately assess margin status with good correlation to gold-standard histopathological examination. Further work, focused on suppressing the optical signal from the monopolar diathermy device, will assist margin assessment on intact WLE specimens and potentially identification of SLN metastases on CLI. SLNB can be performed successfully during ¹⁸F-FDG CLI-guided surgery by using 150 MBq ^{99m}Tc-nanocolloid and blue dye. Based on the results of this study a larger randomized controlled study is warranted to evaluate the impact of intraoperative ¹⁸F-FDG on re-operation rate and Quality of Life in BCS.

DISCLOSURES

David S. Tuch is an employee of and a shareholder in Lightpoint Medical Ltd. The other authors have no conflicts of interest to disclose. This study was supported by funding from Innovate UK, Cancer Research UK King's Health Partners Experimental Cancer Medicine Centre, Guy's and St Thomas' Charity, and the National Institute for Health Research (NIHR) Biomedical Research Centre at Guy's and St Thomas' NHS Foundation Trust and King's College London. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health.

ACKNOWLEDGEMENTS

The authors gratefully acknowledge the excellent support from the King's Health Partners Cancer Biobank, Breast Cancer NOW, Viapath pathology services and the King's College London and Guy's and St Thomas' PET Centre.

REFERENCES

1. St John ER, Al-Khudairi R, Ashrafian H, et al. Diagnostic Accuracy of Intraoperative Techniques for Margin Assessment in Breast Cancer Surgery: A Meta-analysis. *Ann Surg.* 2016.

2. Munshi A, Kakkar S, Bhutani R, Jalali R, Budrukkar A, Dinshaw KA. Factors influencing cosmetic outcome in breast conservation. *Clin Oncol (R Coll Radiol).* 2009;21:285-293.

3. Abe SE, Hill JS, Han Y, et al. Margin re-excision and local recurrence in invasive breast cancer: A cost analysis using a decision tree model. *J Surg Oncol.* 2015;112:443-448.

4. Kalinyak JE, Berg WA, Schilling K, Madsen KS, Narayanan D, Tartar M. Breast cancer detection using high-resolution breast PET compared to whole-body PET or PET/CT. *Eur J Nucl Med Mol Imaging.* 2014;41:260-275.

5. Narayanan D, Madsen KS, Kalinyak JE, Berg WA. Interpretation of positron emission mammography and MRI by experienced breast imaging radiologists: performance and observer reproducibility. *AJR Am J Roentgenol.* 2011;196:971-981.

6. Schilling K, Narayanan D, Kalinyak JE, et al. Positron emission mammography in breast cancer presurgical planning: comparisons with magnetic resonance imaging. *Eur J Nucl Med Mol Imaging*. 2011;38:23-36.

7. Das S, Thorek DL, Grimm J. Cerenkov imaging. *Adv Cancer Res.* 2014;124:213-234.

8. Grootendorst MR, Cariati M, Kothari A, Tuch DS, Purushotham A. Cerenkov luminescence imaging (CLI) for image-guided cancer surgery. *Clinical and Translational Imaging*. 2016:1-14.

9. Mansel RE, MacNeill F, Horgan K, et al. Results of a national training programme in sentinel lymph node biopsy for breast cancer. *Br J Surg.* 2013;100:654-661.

10. The ionising radiations regulations 1999. No. 3232.: Health and Safety Executive; 1999.

11. Ionising radiation (medical exposure) regulations 2000 (IRMER). Department of Health; 2012.

12. The environmental permitting (England and Wales) regulations 2010. Department for environment, food and rural affairs; 2010.

13. Khan KS, Chien PF. Evaluation of a clinical test. I: assessment of reliability. *BJOG.* 2001;108:562-567.

14. Kumar R, Chauhan A, Zhuang H, Chandra P, Schnall M, Alavi A. Clinicopathologic factors associated with false negative FDG-PET in primary breast cancer. *Breast Cancer Res Treat.* 2006;98:267-274.

15. Spinelli AE, Boschi F. Novel biomedical applications of Cerenkov radiation and radioluminescence imaging. *Phys Med.* 2015;31:120-129.

16. Agdeppa ED, Spilker ME. A review of imaging agent development. *AAPS J.* 2009;11:286-299.

17. Spinelli AE, Ferdeghini M, Cavedon C, et al. First human Cerenkography. *J Biomed Opt.* 2013;18:20502.

18. Thorek DL, Riedl CC, Grimm J. Clinical Cerenkov luminescence imaging of (18)F-FDG. *J Nucl Med.* 2014;55:95-98.

19. Hu H, Cao X, Kang F, et al. Feasibility study of novel endoscopic Cerenkov luminescence imaging system in detecting and quantifying gastrointestinal disease: first human results. *Eur Radiol.* 2015;25:1814-1822.

20. Heckathorne E, Dimock C, Dahlbom M. Radiation dose to surgical staff from positron-emitterbased localization and radiosurgery of tumors. *Health Phys.* 2008;95:220-226.

21. Andersen PA, Chakera AH, Klausen TL, et al. Radiation exposure to surgical staff during F-18-FDG-guided cancer surgery. *Eur J Nucl Med Mol Imaging.* 2008;35:624-629.

22. Duran A, Hian SK, Miller DL, Le Heron J, Padovani R, Vano E. Recommendations for occupational radiation protection in interventional cardiology. *Catheter Cardiovasc Interv.* 2013;82:29-42.

23. The 2007 recommendations of the international commission on radiological protection. ICRP publication 103. Ann. ICRP 37 (2-4). International Commission on Radiological Protection; 2007.

24. USNRC. Standards for protection against radiation. 10 CFR 20.1201.

25. Sun JS, Tsuang YH, Chen IJ, Huang WC, Hang YS, Lu FJ. An ultra-weak chemiluminescence study on oxidative stress in rabbits following acute thermal injury. *Burns.* 1998;24:225-231.

26. Loh SA, Carlson GA, Chang EI, Huang E, Palanker D, Gurtner GC. Comparative healing of surgical incisions created by the PEAK PlasmaBlade, conventional electrosurgery, and a scalpel. *Plast Reconstr Surg.* 2009;124:1849-1859.

27. Ruidiaz ME, Cortes-Mateos MJ, Sandoval S, et al. Quantitative comparison of surgical margin histology following excision with traditional electrosurgery and a low-thermal-injury dissection device. *J Surg Oncol.* 2011;104:746-754.

FIGURES WITH LEGENDS



Figure 1: Investigational intraoperative CLI specimen camera. (A) Schematic diagram. Component labels: 1) EMCCD camera, 2) f/.95 lens, 3) Hinged reflex mirror, 4) Complementary metal-oxide-semiconductor reference camera, 5) Specimen table, 6) Lead radiation shielding for EMCCD camera, 7) Focal zone, 8) Fixed lens for reference camera, 9) Filter wheel, 10) LED RGB light array, 11) Specimen chamber. The purple line shows the optical paths for the EMCCD camera and the reference camera as determined by the angle of the reflex mirror. (B) Specimen chamber. The specimen table is placed on a parallelogram to facilitate accurate positioning of the specimen in the center of the image.



Figure 2: WLE specimen from a patient with a grade 3, ER-/HER2-, no special type (NST) carcinoma. (A) Cerenkov image; (B) Grey-scale photographic image overlaid with Cerenkov signal. An increased signal from the tumor is visible (white arrows); mean radiance is 871 ± 131 photons/s/cm²/sr, mean TBR is 3.22. Both surgeons measured the posterior margin (outlined in blue) as 2 mm (small arrow); a cavity shaving would have been performed if the image had been available intraoperatively. The medial margin (outlined in green) measured >5 mm by both surgeons. Pathology ink prevented assessing the lateral margin; a phosphorescent signal is visible (open arrows). (C) Specimen radiography image. The absence of one surgical clip to mark the anterior margin, and the odd position of the superior margin clip (white arrow) prevented reliable margin assessment. (D) Combined histopathology image from two adjacent pathology slides on which the posterior margin (bottom of image) and part of the primary tumor are visible (open arrows). The distance from the posterior margin measured 3 mm microscopically (double arrow). The medial margin is > 5 mm (not present in image).



Figure 3: WLE specimen from a patient with a grade 3, ER+/HER2-, NST carcinoma admixed with high grade DCIS. (A) Cerenkov image; (B) Grey-scale photographic image overlaid with Cerenkov signal. An increased signal from the tumor is visible (white arrows); mean radiance is 406 ± 51 photons/s/cm²/sr, mean TBR is 2.03. The posterior margin (outlined in blue) is 2 mm or 3 mm on CLI as measured by surgeons 1 and surgeon 2, respectively; both surgeons would have performed a cavity shaving. The medial margin (outlined in green) is >5 mm. (C) Specimen radiography image. All 4 radial margins were >5 mm, and both surgeons indicated they would not have performed a cavity shaving. (D) Histopathology image showing the posterior margin (left side of image) and part of the tumor (open arrows). The posterior margin was 5 mm distant histologically (double arrow).



Figure 4: WLE specimen from a patient with a grade 3, ER+/HER2-, NST carcinoma admixed with high grade DCIS. (A) Cerenkov image; (B) Grey-scale photographic image overlaid with Cerenkov signal. An elevated signal (white arrow) from the tumor can be seen. Mean radiance is 637 ± 47 photons/s/cm²/sr; mean TBR is 1.63. Both surgeons measured the posterior margin (outlined in blue), medial margin (outlined in green) and lateral margin (outlined in red) distances as >5 mm; a cavity shaving would not have been performed based on the CLI image. (C) Specimen radiography image. All 4 radial margins were >5 mm as measured by both surgeons and did not prompt resection of cavity shave margins (D) Histopathology image from large-format pathology block. The tumor is >5 mm from the posterior margin (solid arrow), medial margin (dashed arrow) and lateral margin (not visible in image).

Patient	Tumor type ¹	Histologic grade invasive (1 – 3)	ER/HER2 status of invasive cancer (Pos/Neg) ²	Mean tumor radiance (photons/s/cm2/sr)	TBR ³	Margin distance CLI surgeon 1, 2 ⁷ (mm)	Margin distance histopath ⁵ (mm)	Tumor size CLI surgeon 1, 2 ⁷ (mm)	Invasive tumor size histopath ⁸ (mm)	Whole tumor size histopath ⁸ (mm)
1	NST DCIS	2	Pos/Neg	_4	-	-	-	-	13	13
2	NST DCIS	3	Neg/Pos	453.59	2.34	6, 6	>5	*6	22	22
3	NST	3	Neg/Neg	871.16	3.22	2, 2 28, 30	3 >5	20, 18	20	20
4	NST DCIS	3	Neg/Neg	_4	-	-	-	-	14	14
5	NST DCIS	3	Pos/Neg	405.76	2.03	2, 3	5	18, 18	20	20
6	ILC	2	Pos/Neg	544.04	2.44	9, 9	>5	20, 19	22	22
7	NST	3	Pos/Neg	667.47	2.72	10, 11	>5	*6	25	25
8	NST DCIS	2	Pos/Neg	308.30	1.63	6, 8 16, 15	>5 >5	19,19	15	35
9	NST DCIS	3	Pos/Neg	593.93	3.08	14, 13 6, 7	>5 >5	14, 15	18	19
10	NST DCIS	3	Pos/Neg	648.29	2.46	8, 8	>5	22, 22	19	29
11	NST DCIS	2	Pos/Neg	466.03	2.54	15, 14	>5	13, 11	13	13
12	NST DCIS	3	Pos/Neg	637.08	1.63	9, 9 30, 31 22, 22	>5 >5 >5	12, 10	14	14

Table 1: CLI and postoperative histopathology results for each patient in the analysis dataset.

1. NST: invasive carcinoma of ductal/no special type, DCIS: ductal carcinoma in situ, ILC: invasive lobular carcinoma

2. Pos: positive, Neg: negative

3. Tumor-to-background ratio

4. No elevated tumor radiance on CLI

5. Histopath: histopathology

6. Presence of orientation inks prevented measuring tumor size on CLI.

7. Margin distance and tumor size are shown for surgeon 1 and 2, respectively

8. The histopathological tumor size displayed in the table is the tumor size measured in the same direction as the tumor size measurement on CLI. In patient 3, 11 and 12 the largest invasive and whole tumor size (i.e. extent of DCIS and invasive cancer) was measured in a different direction, and was 32 mm, 33 mm and 12 mm, respectively

Table 2: Measured effective radiation doses by occupation from 22 surgical procedures.

Staff Group	N^1	Mean Effective Dose	Range (µSv)	Estimated number of procedures per individual per year ²	Estimated number of procedures per individual per year ²	
		per procedure ± std (µSv)		ICRP ³ (20 mSv annual limit)	USNRC ⁴ (50 mSv annual limit)	
Surgeon	46	34 ± 15	8 – 74	270	676	
Anesthetist	22	11 ± 5	0 – 18	1111	2778	
NM technologist	22	9 ± 4	1 – 15	1333	3333	
Anesthetist assistant	22	6 ± 3	0 – 11	1818	4545	
Trial Co-coordinator	21	5 ± 2	1 – 10	2000	5000	
Recovery Nurse	43	4 ± 3	0 – 14	1429	3571	
Scrub Nurse	22	2 ± 1	0 – 5	4000	10000	
Periphery Nurse	23	1 ± 1	0 – 4	5000	12500	
Research fellow	36	1 ± 2	0 – 13	1538	3846	
Ward Nurse	15	0	0 – 1	20000	50000	
Tissue Biobank Practitioner	14	0	0 – 1	20000	50000	

1. N = number of measurements

Based on maximum Effective Dose per procedure per staff group
 ICRP: International Commission on Radiological Protection
 USNRC: United States Nuclear Regulatory Commission