



King's Research Portal

DOI:

[10.1016/j.jcin.2016.11.048](https://doi.org/10.1016/j.jcin.2016.11.048)

Document Version

Peer reviewed version

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

De Silva, K., Roy, J., Webb, I., Dworakowski, R., Melikian, N., Byrne, J., MacCarthy, P., & Hill, J. (2017). A Calcific, Undilatable Stenosis: Lithoplasty, a New Tool in the Box? *Jacc-Cardiovascular Interventions*, *10*(3), 304-306. <https://doi.org/10.1016/j.jcin.2016.11.048>

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

**A calcific, undilatable stenosis;
Lithoplasty – a new tool in the box?**

Kalpa De Silva MBBS PhD, James Roy MD, Ian Webb MA PhD, Rafal Dworakowski MD PhD,
Narbeh Melikian MBBS PhD, Jonathan Byrne MBBS PhD, Philip MacCarthy BSc MBBS PhD,
Jonathan Hill MA PhD

Affiliation: Cardiology Department, King's College Hospital NHS Trust, London, UK

Address for Corresponding Author:

Dr Jonathan Hill

King's College Hospital
Denmark Hill
London, SE5 9RS, UK

Email: jmhill@nhs.net

Tel: +447739 988 822

Fax: +44203 299 3445

Submission for Images in Intervention

Word count: 312

There has been no financial support, funding or relevant relationships with industry for any of the listed authors.

A 69-year-old gentleman with established coronary artery disease and LV dysfunction (EF 40%) with typical CCS class 3 angina underwent PCI for severe diffuse calcific disease in the RCA (Fig 1). The vessel was prepared with 2.5mm and 3.0mm balloon pre-dilatation. However, despite the use of 3.0mm non-compliant balloon there was inadequate balloon expansion with a 'dog-bone' appearance observed (Fig 2). The patient was readmitted for a further attempt at PCI with adjunctive lithoplasty for calcium debulking. A 3.5x12mm Lithoplasty balloon (Shockwave Medical) was the initial and only balloon used to pre-dilate and treat the entire length of disease. Briefly, the Lithoplasty technique involves inflating the balloon to low pressure (4atm) with eight, 10-second pulses of ultrasound energy delivered per balloon. A degree of balloon deformation was observed secondary to a region of lesion constriction, at the nominal pressure of 4atm, the Lithoplasty was initiated, with the balloon seen to inflate fully at this low pressure (Movie 1). The balloon is then inflated to 6atm for 15-20 seconds after each pulse, to maximise balloon expansion and aid removal of debris. Lithoplasty preferentially allows calcium modification without affecting the endovascular soft tissue, and subsequently aids stent delivery and optimisation. OCT performed pre and post-lithoplasty showed the calcium 'cracking' effect of the technique (Fig 3). The segment of disease was then treated with a 3.5x22 and 4.0x16mm DES with a good OCT (Fig 4) and angiographic (Fig 5 & Movie 2) result.

Calcific coronary disease remains an important cause of stent under expansion and represents a nidus for stent thrombosis. Revascularization undertaken in those of > 75 years, a cohort with increased coronary calcification, now accounts for 25-30% of all PCI procedures.¹ Current techniques to modify calcific stenoses include, standard or high-pressure non-compliant balloons, cutting/scoring balloons or rotational atherectomy. High pressure balloon treatment may lead to localized wall injury, which may provide a vascular substrate for restenosis, with this and rotational atherectomy also increasing the risk of coronary perforation (0.7% with rotational atherectomy versus 0.1% in standard procedures).² Lithoplasty may provide an adjunct to PCI that provides focal calcium modification with limited localized injury to the endovascular surface whilst aiding stent delivery and expansion.

References:

¹Vandermolen S, Abbott J, De Silva K. What's Age Got to Do with it? A Review of Contemporary Revascularization in the Elderly. *Curr Cardiol Rev.* 2015;11(3):199-208.

²Cohen BM, Weber VJ, Reslman M, Casale A, Dorros G. Coronary perforation complicating rotational atherectomy: the U.S. Multicenter experience. *Cathet Cardiovasc Diag,* 1996;Suppl 3:55-9.

Figures

Figure 1: Pre-PCI Angiogram

Figure 2: 'Dog-bone' of non-compliant balloon

Figure 3: OCT showing 'calcium cracking' following Lithoplasty

Figure 4: Co-registered angiogram/OCT result post PCI

Figure 5: Final angiographic appearance

Movie 1: Zoomed view of Shockwave balloon at 4atm during Lithotripsy pulse

Movie 2: OCT imaging post PCI; Short-axis and Longitudinal views with rendered stent and intra-coronary 'fly-through'